

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 10/27/99

Date Received: 10/14/99

Project: Metro Self Monitor, PO# M62806

Date Samples Extracted: 10/18/99

Date Extracts Analyzed: 10/21/99

**RESULTS FROM THE ANALYSIS OF THE WATER SAMPLE
FOR CHROMIUM, COPPER, NICKEL, ZINC
USING METHOD 6010**

Samples Processed Using Method 3005A

Results Reported as mg/L (ppm)

<u>Sample ID</u>	<u>Chromium</u>	<u>Copper</u>	<u>Nickel</u>	<u>Zinc</u>
M62806	0.43	0.27	0.36	<0.05
Method Blank	<0.05	<0.05	<0.05	<0.05

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 10/27/99

Date Received: 10/14/99

Project: Metro Self Monitor, PO# M62806

**QUALITY ASSURANCE RESULTS
FOR TOTAL METALS BY
INDUCTIVELY COUPLED PLASMA (ICP)
(METHOD 6010)**

Laboratory Code: 910081-01 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	mg/L (ppm)	0.43	0.35	19	0-20
Copper	mg/L (ppm)	0.27	0.25	8	0-20
Nickel	mg/L (ppm)	0.36	0.28	21 vo	0-20
Zinc	mg/L (ppm)	<0.05	<0.05	nm	0-20

Laboratory Code: 910081-01 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	% Recovery MS	% Recovery MSD	Acceptance Criteria	RPD
Chromium	mg/L (ppm)	5	0.43	96	94	80-120	2
Copper	mg/L (ppm)	5	0.27	108	108	80-120	0
Nickel	mg/L (ppm)	10	0.36	105	106	80-120	1
Zinc	mg/L (ppm)	5	<0.05	116	106	80-120	9

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	% Recovery LCS	% Recovery LCSD	Acceptance Criteria	RPD
Chromium	mg/L (ppm)	5	117	120	80-120	3
Copper	mg/L (ppm)	5	106	108	80-120	2
Nickel	mg/L (ppm)	10	100	102	80-120	2
Zinc	mg/L (ppm)	5	101	103	80-120	2

nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.

vo - The value reported fell outside the control limits established for this analyte.

(DH 10/14/89 AI,

SAMPLE CHAIN OF CUSTODY

Company AMERICAN Copper Works Contact Gerald A. Thompson
Address 628 Hawthorn St
City, State, Zip Seattle WA 98134
Phone # 206-382-8379 FAX # 206-382-4309 Date 10-14-99

PURCHASE ORDER #

in 62806



PROJECT LOCATION

3200 6th Ave S.

SAMPLE DISPOSAL INFORMATION

100

Dispose after 30 days
Return Samples
Call for Instructions

SIGNATURE	PRINT NAME	COMPANY	Date	Time
Relinquished by: 	Gerald Thompson	ACW.	10-8-88	12:13pm
Received by: 	DIANNE C. HORK	FBI	10/14/88	12:13pm
Relinquished by:				
Received by:				

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Jensen, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

October 27, 1999

DUPLICATE COPY

INVOICE #99ACU1027-1

Accounts Payable
Alaskan Copper Works
628 South Hanford St.
Seattle, WA 98134

RE: Project Metro Self Monitor, PO# M62806: Results of testing requested
by Gerry Thompson for material submitted on October 14, 1999.

1 water sample analyzed for
Chromium, Copper, Nickel, and Zinc
using Method 6010 @ \$80 per sample

\$ 80.00

Amount Due

\$ 80.00

PLEASE NOTE: As a convenience for our valued clients, we can accept credit cards (VISA, MasterCard or government purchasing cards) for payment on our invoices. Please contact our office if you would like to take advantage of this service.

FEDERAL TAX ID (b) (6)

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Jensen, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

October 27, 1999

Gerry Thompson, Project Manager
Alaskan Copper Works
628 South Hanford St.
Seattle, WA 98134


Dear Mr. Thompson:

Included are the results from the testing of material submitted on October 14, 1999 from your Metro Self Monitor, PO# M62806 project. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Dianne C. Hurt
Project Manager

Enclosures
ACU1027R.DOC